



4 November 2021

(21-8362)

Page: 1/2

Committee on Technical Barriers to Trade

Original: Spanish

NOTIFICATION

The following notification is being circulated in accordance with Article 10.6.

1. Notifying Member: <u>PERU</u> If applicable, name of local government involved (Articles 3.2 and 7.2):
2. Agency responsible: <i>Instituto Nacional de Calidad - Dirección de Metrología</i> (National Institute for Quality - Directorate of Metrology) Calle las Camelias 817, San Isidro - Lima 27 - Peru Tel.: +(51-1) 640-8820, Ext. 1516 Email: jarce@inacal.gob.pe Name and address (including telephone and fax numbers, email and website addresses, if available) of agency or authority designated to handle comments regarding the notification shall be indicated if different from above: <i>Ministerio de Comercio Exterior y Turismo, MINCETUR</i> (Ministry of Foreign Trade and Tourism) Calle Uno Oeste Nº 50 - Urb. Corpac - Lima 27 - Peru Tel.: +(51-1) 513-6100, Ext. 1223 or 1239 Email: otc@mincetur.gob.pe
3. Notified under Article 2.9.2 [X], 2.10.1 [], 5.6.2 [], 5.7.1 [], other:
4. Products covered (HS or CCCN where applicable, otherwise national tariff heading. ICS numbers may be provided in addition, where applicable): Other ophthalmic instruments and appliances (HS 901850); Ophthalmic equipment (ICS 11.040.70)
5. Title, number of pages and language(s) of the notified document: <i>Proyecto de Norma Metrológica Peruana PNMP 02X:202X "INSTRUMENTOS OFTÁLMICOS - TONÓMETROS DE IMPRESIÓN Y APLANAMIENTO - Parte 1: Requisitos técnicos y metrológicos; Parte 2: Procedimientos de pruebas; y Parte 3: Formato de informe de pruebas" - 1ra Edición</i> (Draft Peruvian Metrology Standard (PNMP) No. 02X:202X: "Ophthalmic instruments - Impression and applanation tonometers - Part 1: Metrological and technical requirements; Part 2: Test procedures; and Part 3: Test report format" - 1st Edition) (48 page(s), in Spanish)
6. Description of content: The notified draft Peruvian Metrology Standard specifies requirements for the design compliance and construction of impression and applanation tonometers used for the determination of intraocular pressure (IOP) in clinical applications. When referring to applanation tonometers, which measure the force necessary to applanate the cornea over a given diameter using the flat contact side of a pressure body, the notified Standard covers the method developed by Hans Goldmann.
7. Objective and rationale, including the nature of urgent problems where applicable: The notified text explains the rationale for the preparation of draft Peruvian Metrology Standard (PNMP) No. 02X:202X: "Ophthalmic instruments - Impression and

	applanation tonometers - Parts 1, 2 and 3 - 1st Edition". The draft Standard was prepared by means of an "adoption system" for the preparation of Peruvian Metrology Standards, in accordance with Article 9(A) of the Procedure for the Drafting and Approval of Peruvian Metrology Standards - 2nd Edition, as approved by Directorial Resolution No. 001-2021-INACAL/DM, published on 13 January 2021; Others
8. Relevant documents: International Recommendation OIML R 145, Edition 2015 (E): "Ophthalmic instruments - Impression and applanation tonometers. Part 1: Metrological and technical requirements; Part 2: Test procedures; and Part 3: Test report format".	
9. Proposed date of adoption: To be determined. Proposed date of entry into force: Twelve (12) months after publication in the Official Journal <i>El Peruano</i> .	
10. Final date for comments: Sixty (60) days from notification.	
11. Texts available from: National enquiry point [] or address, telephone and fax numbers and email and website addresses, if available, of other body: <i>Ministerio de Comercio Exterior y Turismo, MINCETUR</i> (Ministry of Foreign Trade and Tourism) Calle Uno Oeste Nº 50 - Urb. Corpac - Lima 27 - Peru Tel.: +(51-1) 513-6100, Ext. 1223 or 1239 Email: otc@mincetur.gob.pe <i>Instituto Nacional de Calidad - Dirección de Metrología</i> (National Institute for Quality - Directorate of Metrology) Calle las Camelias 817, San Isidro - Lima 27 - Peru Tel.: +(51-1) 640-8820, Ext. 1516 Email: jarce@inacal.gob.pe https://www.inacal.gob.pe/metrologia/categoria/normasmetrologicas http://extranet.comunidadandina.org/sirt/public/buscapalavra.aspx http://consultasenlinea.mincetur.gob.pe/notificaciones/Publico/FrmBuscador.aspx https://members.wto.org/crnattachments/2021/TBT/PER/21_6995_00_s.pdf https://members.wto.org/crnattachments/2021/TBT/PER/21_6995_01_s.pdf https://members.wto.org/crnattachments/2021/TBT/PER/21_6995_02_s.pdf	